

Appropriate UK enforcing authorities

To submit a notification, the appropriate IR(ME)R enforcing authorities are:

England:

[Care Quality Commission: IR\(ME\)R notification](#)

Wales:

[Healthcare Inspectorate Wales](#)

email: IRMERIncidents@Wales.GSI.Gov.uk

Northern Ireland:

[The Regulation and Quality Improvement Authority](#)

Scotland:

[Healthcare Improvement Scotland](#)

email: hcis.irmer@nhs.net

Reporting device-related incidents

Where there are risks to individuals relating to medical devices, employers should consider reporting all device and medicine-related incidents to other agencies including:

England and Wales:

[The Medicines and Healthcare products Regulatory Agency \(MHRA\)](#)

Scotland:

[Health Facilities Scotland Incident Reporting and Investigating Centre \(IRIC\)](#)

email: nss.irc@nhs.scot

Northern Ireland:

[The Northern Ireland Adverse Incident Centre](#)

It is good practice for employers to report this type of incident (even if they have not resulted in a SAUE). This enables the UK Competent Authority for the Medicines and Medical Device Regulations (MHRA) to take appropriate action with the manufacturer.

Public or occupational exposures

Where a member of the public or a worker receives an over-exposure to ionising radiation, this needs to be reported to the [Health and Safety Executive](#) under Regulation 26 of The Ionising Radiation Regulations 2017.

Over-exposures resulting from equipment faults before the equipment is put into clinical use, for example for critical examination, should also be reported to the Health and Safety Executive.

[Health and Safety Executive: Ionising radiation](#)

Health and Safety Executive Northern Ireland

[Health and Safety Executive Northern Ireland: Ionising radiation](#)

© Care Quality Commission